Daily Checklist for TABS Pressure Pads

Activation Date: _____

(*Date the pressure pad is put into use.*)

Serial # _____

Testing the System

Test the TABS monitoring system before each use and inspect the pressure pads and the TABS monitor regularly to make sure they are not damaged. Test the pressure pads by pressing firmly on the pressure pad for 3 seconds. When you remove the pressure, the alarm will sound. Reset the alarm.

Date	Shift	Initials	Comments

For assistance, please call Stanley-Senior Technologies: 800-824-4490 or 402-475-4002.

Follow the installation and testing instructions outlined in the TABS Installation Instruction Manual.

TABS is a registered trademark of The Stanley Works © 2003. DOC-30-20010-AA