

M200 Approved Cleaning Materials

This Technical Bulletin provides important information about which cleaning materials have been approved by STANLEY Healthcare for use with the M200/ M200-C Fall Monitors.



Approved Cleaning Materials

Cleaning Material (Approved)	
A-456-N (Airkem)	SaniWipes / Purple Top
Asepti-zyme	SaniWipes / Red Top = Sani Cloth Plus
Bleach used with gauze (Clorox germicidal bleach)	Spor Klenz - ready to use
Cavi Cide or Cavi Wipes	Sporicidin®
CaviWipes (Metrex)	Staph-Attack
CLOROX Commercial Solutions® Germicidal Wipes	Thymocide
Coverage™ HBV	Thymocidine
Detergezime - generic = detergent + enzyme	Vesphene® II SE
Expose 256	Virex II 256
Expose II 256 Or Phenolic 256 DC	Virkon -S Powder
Formula C™	Virox Accel Prevention (wipes & concentrate in use)
Frekaderm Actief 100 g liquid	Virox Rescue
HB Quat Disinfectant Cleaner Concentrate 25 (3M)	Wex Cide 128
Isopropyl Alcohol used with gauze	Wex Cide-Ready-To-Use
LpH Disinfectant Cleaner	Wexide
Maxima 128	
Metrizyme	
Neutral QUAT Disinfectant Cleaner Concentrate	
OPTIPRO SOLID MULTI ENZYMATIC MANUAL CLEANER	
Precise	
Premoistened Alcohol/DI Clean Wipes	
Professional Amphyl Hospital Bulk Disinfectant	
Prolystica Ultra Concentrate Enzymatic Cleaner	
Sani-Cloth AF3 Germicidal Disposable Wipes	

About STANLEY Healthcare

Over 15,000 hospitals and senior living communities rely on STANLEY Healthcare solutions to empower caregivers to deliver better care. STANLEY Healthcare is a part of Stanley Black & Decker and a proud supporter of the Alzheimer's Association®. Learn more at www.stanleyhealthcare.com.